

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32312-5K	2024-05-03
Low	32312-51	2024-05-03
Normal	32312-52	2024-05-04
High	32312-53	2024-05-10

SWELAB

		CONTROL	L
		LOT	32312-51
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.24	0.12
MCV	fL	65.2	7.0
RDW	%	14.0	4.5
Hct	%	14.6	3.0
Plt	10 ⁹ /L	85	15
MPV	fL	12.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.4	0.3
MCH	pg	24.1	2.0
MCHC	g/dL	37.0	4.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.9	7.0
Mid	%	3.0	5.0
Gran	%	49.1	7.0

		CONTROL	N
		LOT	32312-52
		Mean ²	± Range ³
		4.20	0.18
		80.4	7.0
		14.4	4.5
		33.8	3.5
		244	30
		12.3	1.5
		8.7	0.6
		12.0	0.4
		28.6	2.0
		35.5	4.0
		4.4	0.6
		0.3	0.5
		4.0	1.0
		50.7	5.0
		3.2	5.0
		46.1	5.0

		CONTROL	H
		LOT	32312-53
		Mean ²	± Range ³
		5.12	0.22
		92.1	7.0
		12.4	4.5
		47.2	4.0
		538	60
		12.5	1.5
		20.7	1.8
		16.3	0.5
		31.8	2.0
		34.5	4.0
		13.3	1.5
		0.6	1.0
		6.8	2.5
		64.1	5.0
		2.9	5.0
		33.0	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dny (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nožimė
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

25122-4



www.swelab.com/
support/download



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 32312-51
Prod.: Con-Diff Low
ID: 3231251+
Exp.: 3-May-2024

Instr.: Swelab Alfa
Lot: 32312-52
Prod.: Con-Diff Norm
ID: 3231252+
Exp.: 4-May-2024

Instr.: Swelab Alfa
Lot: 32312-53
Prod.: Con-Diff High
ID: 3231253+
Exp.: 10-May-2024

